

HELAGO-CZ, s.r.o.
Commercial Register maintained by the Regional Court in Hradec Králové
Section C, File 17879
Kladská 1082
500 03 Hradec Králové 3
Company ID: 25 96 39 61, VAT: CZ 25963961

Phone: 495 220 229, 495 220 394

Fax: 495 220 154 GSM gate: 602 123 096 E-mail: info@helago-cz.cz Web: http://www.helago-cz.cz

Airway Assessment Training Model

Order code: 4119.BTCSIE



Information about product price on demand

Parameters

Quantitative unit ks

- Objective intubation practice with an adult torso model
- Ability to do realistic practice by using skin-like feeling Human torso model Head tilt / Chin lift, Sniffing position
- Jaw Thrust maneuver (Displayed through the monitor)
- Ventilation by using Bag-Valve-Mask
- Insert the oral and nasal airway can be applied
- Intubation tube (ETT), laryngeal mask (LMA/I-gel), combi tube (Combi-Tube) can be applied
- Intubation by using laryngoscope
- Observable chest rising during the ventilation and check with stethoscope
- Alarm sounds due to the excessive use of laryngoscope
- Detects Head tilt-chin lift method, Jaw Trust, sniffing position and display these through a tablet
- Observable intubation depth (deep/moderate/shallow) through tablet
- Observable esophageal intubation on the tablet PC
- Detects tidal volume and stomach expansion and display the state on the Tablet PC
- User can check up the training record through the auto-ID items and check-list
- Check-list & result-point is all configurable.
- All the results can be saved.
- Wireless Blue-tooth connection between display device-tablet PC

- Adult Torso Model (Controller(installed) / Blue-tooth device(installed))
- Lubricant
- AC Power Adaptor
- Battery Holder
- Storage Case

Tablet PC (Software include) - option