

HELAGO-CZ, s.r.o. Commercial Register maintained by the Regional Court in Hradec Králové Section C, File 17879 Kladská 1082 500 03 Hradec Králové 3

Company ID: 25 96 39 61, VAT: CZ 25963961 Phone: 495 220 229, 495 220 394

Fax: 495 220 154 GSM gate: 602 123 096 E-mail: info@helago-cz.cz Web: http://www.helago-cz.cz

## DT514 - Ultrasound bath DT 514

Order code: 2007.3250



Information about product price on demand

**Parameters** 

Quantitative unit ks

## **Description**

- full convenience in use and display for all bath types
- easy to operate self-explanatory symbols well recognizable parameters,
- target as well as actual values brightly lighting LED-display
- 4-step-operation as "plug and clean"
- heater 20-80 °C, thermostatically adjustable, with LED-display for target value and actual value of temperature
- optical signal in case of excess temperature
- timer for countdown and continuous operation
- DEGAS: adjustable pulse sound level for removing unwanted air bubbles, visible at LCD-display
- automatic short-time ultrasound during heating
- up period to avoid delay in boiling switch-off possible
- CE marked as medical device
- quality management system according to the requirements of EN ISO 13485/11.2000 in the production line
- protection class IP 33 at highest possible level through liquid-tight mains cable and housing front without openings
- SweepTec for permanent sound field oscillation ensures even and gentle cleaning
- utility patent DE 202 21 041
- automatical safety shut-down 12 hours after last keypress
- marking of filling level for safe dosage
- one-piece welded drain bow
- PZT-transducers with high efficiency

For more information about this product please download product sheet under "Files for download".