

HELAGO-CZ, s.r.o. Commercial Register maintained by the Regional Court in Hradec Králové Section C, File 17879 Kladská 1082 500 03 Hradec Králové 3

Company ID: 25 96 39 61, VAT: CZ 25963961

Phone: 495 220 229, 495 220 394 Fax: 495 220 154 GSM gate: 602 123 096

E-mail: info@helago-cz.cz Web: http://www.helago-cz.cz

DL102H - Ultrasound bath DL 102

Н

Order code: 2007.7180



Information about product price on demand

Parameters

Quantitative unit ks

Description

- Cleaning of sensitive materials
- Mixing of samples
- Homogenizing of samples for analysis
- Fast degassing of solvents for HPLC

Features

- Power control from 20 to 100 %
- Heating from 20 to 80 °C, thermostatically adjustable, LED display
- Optical LED signal in case of excess temperature
- Automatic short-time extra ultrasound during heating up period to avoid delay in boiling switch-off possible
- Timer for countdown or continuous operation
- DEGAS quick degassing
- SweepTec® permanent sound field oscillation ensures even and gentle cleaning
- Automatical safety shut-down 12 hours after last keypress
- Marking of filling level for safe dosage

- Drain valve for simple emptying, one-piece welded drain bow
- Protection class IP 33 at highest possible level through liquid-tight mains cable and housing front without openings

For more information about this product please download product sheet under "Files for download".