



HELAGO-CZ, s.r.o.
Commercial Register maintained by the Regional
Court in Hradec Králové
Section C, File 17879
Kladská 1082
500 03 Hradec Králové 3
Company ID: 25 96 39 61, VAT: CZ 25963961
Phone: 495 220 229, 495 220 394
Fax: 495 220 154
GSM gate: 602 123 096
E-mail: info@helago-cz.cz
Web: <http://www.helago-cz.cz>

Hand Refractometer HR-160N for Urine and Protein, ATC, Built-in LED Illuminator

Order code: **0504.HR160N**



Cena bez DPH

59,60 Eur

Price with VAT

72,12 Eur

Parameters

Types of refractometers

Hand refractometers

What do I measure

Refractive Index RI, Urine, Protein

Serie

HR-N

Scale

g/dl | sg RI

Range	0-12 1.000 - 1.050 1.330 - 1.360
Description	for Urine and Protein, ATC, Built-in LED Illuminator
Graduation	0.2 0.005 0.0005RI
Display	Analog
Quantitative unit	ks

Hand Refractometer for Urine and Protein, ATC, Built-in LED Illuminator

Measuring range:

- Serum Protein 0-12g/dl
- Specific Gravity 1.000-1.050sg
- Refractive Index: 1.330RI-1.360RI.

Resolution:

- Serum Protein 0.2g/dl
- Specific Gravity 0.005sg
- Refractive Index: 0.0005RI.