

HELAGO-CZ, s.r.o. Commercial Register maintained by the Regional Court in Hradec Králové Section C, File 17879 Kladská 1082 500 03 Hradec Králové 3 Company ID: 25 96 39 61, VAT: CZ 25963961

Phone: 495 220 229, 495 220 394 Fax: 495 220 154 GSM gate: 602 123 096

E-mail: info@helago-cz.cz Web: http://www.helago-cz.cz

## Stericel 111 ECO - Hot Air Sterilizer Stericell, volume 111 l

Order code: 2102.010572222



Information about product price on demand

## **Parameters**

Volume [I] 111

Inner dimensions (w × h × d) [mm]  $540 \times 530 \times 370$ 

Quantitative unit

- Hot-air sterilizer meets the criteria of a medical device
- 111-liter chamber volume
- Working temperature of from +10°C over ambient temperature up to 250°C
- Noiseless operation with a fine patented system of forced air flow in the chamber thanks to a built-in fan eliminating "cold air" area formation

ks

- High operation comfort; precise temperature and time regulation; three preset sterilization programs
- Suitable for hot-air sterilization at health centers for laboratory purposes; potential use as an oven

- Model with a Fuzzy-Logic microprocessor control system; control elements situated on a foil keyboard, process information indicated on a LED display
- Wide range of optional equipment and accessories
- Output for a printer or PC connection
- Validation possibility (IQ , OQ)