



HELAGO-CZ, s.r.o.  
Commercial Register maintained by the Regional Court in  
Hradec Králové  
Section C, File 17879  
Kladská 1082  
500 03 Hradec Králové 3  
Company ID: 25 96 39 61, VAT: CZ 25963961  
Phone: 495 220 229, 495 220 394  
Fax: 495 220 154  
GSM gate: 602 123 096  
E-mail: [info@helago-cz.cz](mailto:info@helago-cz.cz)  
Web: <http://www.helago-cz.cz>

**VET14020 - Dog Abdomen - Phantom  
for CT and X-ray**  
Order code: **4003.VET14020**



Information about product price on demand

Parameters

Quantitative unit

ks

The phantom provides a realistic simulation of a dog abdomen with no contrast agent. Bones and soft tissue are displayed authentically with realistic CT values for all tissues at 120 kVp tube voltage in the CT. Air spaces are filled with a cellulose polymer composite with about -80 Hounsfield units. If the phantom is to be used primarily with other tube voltages (e.g. 100 kVp), the calibration of the CT values can be adjusted accordingly if required. The phantom also provides realistic tissue contrasts in X-ray imaging.

Size: 19x18x20cm