



HELAGO-CZ, s.r.o.  
Commercial Register maintained by the Regional Court in Hradec Králové  
Section C, File 17879  
Kladská 1082  
500 03 Hradec Králové 3  
Company ID: 25 96 39 61, VAT: CZ 25963961  
Phone: 495 220 229, 495 220 394  
Fax: 495 220 154  
GSM gate: 602 123 096  
E-mail: [info@helago-cz.cz](mailto:info@helago-cz.cz)  
Web: <http://www.helago-cz.cz>

## MBB-12 - Laminar box biohazard

Order code: **1501.MBB12**



Information about product price on demand

Parameters

Inner dimensions (w × d × h) [mm]	1225 × 580 × 690
Ext. dimensions (w×d×h)	1230 × 770 × 1400
Quantitative unit	ks

Microbiological safety cabinet (MSC, BSC), class II. according to the standard EN 12469: 2000, it is intended for applications requiring laminar air flow to protect the product from particulate and bacterial contamination and at the same time requiring protection of the operator and the environment from the influence of the processed product. The MBB workstation is a new generation device, equipped with the latest components, security elements and a SMART control system.

- **Fully automated operation controlled** by a smart control system CR2000 with a 7 "touch panel, ensuring optimal parameters in the workspace for the selected operating mode, maximum security, remote control, advanced timers, accessory control, email and mobile phone notifications, and much more
- **Front glass with electric movement** with automatic positioning and advantage protection against injury and damage
- **Ergonomic design, large workspace**
- **Quiet operation** thanks to perfect design, good sound insulation, powerful motor with low noise. The operator can therefore concentrate on working for a longer period of time
- **Low vibration** is achieved by choosing low vibration components, anti-vibration components and special, flexible, fasteners and hinges. Thanks to this, the workstation is also suitable for vibration-sensitive operations, such as working with a microscope, weighing, etc.
- **Continuous monitoring** of main critical parameters such as filter clogging, laminar airflow rate in the working space and air flow at the outlet of the MSC, temperature, etc.
- **Early warning system** for filter replacement, qualification, service, validation, etc.
- **Hour meter** for UV-C lamp, filters, fan
- **Audit trail** registration of setting changes, alarms and other information
- **Temperature sensor** in the working space